

### QC Analysis Report

Test Name	Reported Result	Units	%R or RPD	QC Flag	Low Limit	Upper Limit
Sodium	2.1	mg/L	103		85	115
Chlorodibromomethane	19	%	94		80	120
Chloroform	20	%	100		80	120
Bromoform	19.7	%	98		80	120
Chloroform	20.8	%	104		80	120
Dibromochloromethane	20	%	100		80	120
Calcium	2.1	mg/L	104		85	115
Bromoform	18	%	88		80	120
Potassium	1.9	mg/L	97		85	115
Bromate	0.25	%	100		80	120
Sulfate	11	%	110		80	120
Chloride	5.23	%	105		80	120
Magnesium	1.9	mg/L	97		85	115
Alkalinity	360	mg/L	3		0	20
Chlorine	0.24	%	4		0	20
Turbidity	1.31	%	2		0	20
Sulfate	2100	%	5		0	20
Chloride	13.1	%	4		0	20